
Midmark M11 - v1.0.0
Total Cycles: 4935

Name: RITTER 1
Sterilizer ID: V2178198
09/24/2024 12:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 98.5 F 0.0
2:00 99.5 F 0.0
4:00 121.6 F 0.1
6:00 156.2 F 0.2
8:00 183.3 F 0.4
10:00 205.3 F 2.8
12:00 216.8 F 6.1
14:00 225.5 F 9.7
16:00 237.1 F 10.7
18:00 237.1 F 10.3
20:00 235.4 F 8.3
22:00 246.3 F 13.5
24:00 244.1 F 12.1
26:00 256.3 F 18.5
28:00 267.3 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.6 F 28.7
2:00 271.6 F 27.9
3:00 272.2 F 28.7
4:00 272.4 F 28.5

Min 271.5 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.9 F 0.0
5:00 210.1 F 0.0
10:00 201.1 F 0.0
15:00 203.8 F 0.0
20:00 205.5 F 0.0
25:00 208.7 F 0.0
30:00 210.3 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 29:29
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:07:50
CYCLE COMPLETE

09/24/2024 02:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____