

-----  
Midmark M11 - v1.0.5  
Total Cycles: 675

Name: RITTER2  
Sterilizer ID: V2313228  
01/10/2022 11:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.6 F 0.0  
2:00 161.4 F 2.2  
4:00 165.4 F 5.5  
6:00 189.6 F 10.0  
8:00 209.9 F 5.4  
10:00 224.7 F 8.5  
12:00 236.0 F 13.6  
14:00 238.8 F 11.2  
16:00 237.8 F 9.8  
18:00 251.5 F 16.5  
20:00 263.1 F 23.3  
22:00 272.1 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.9  
1:00 272.5 F 29.5  
2:00 271.4 F 28.9  
3:00 272.6 F 29.6  
4:00 271.8 F 28.9

Min 271.4 F 28.8  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.7 F 0.0  
5:00 214.5 F 0.0  
10:00 215.7 F 0.0  
15:00 209.2 F 0.0  
20:00 203.4 F 0.0  
25:00 213.5 F 0.0  
30:00 229.3 F 0.0  
Drying Complete

FILLING: 1:31  
HEATING: 22:11  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 01:00:44  
CYCLE COMPLETE

01/10/2022 12:07 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_