

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3574

Name: RITTER 1  
Sterilizer ID: V2178198  
07/24/2023 07:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 190.1 F 0.0  
2:00 190.7 F 3.4  
4:00 198.7 F 8.9  
6:00 222.6 F 6.8  
8:00 230.9 F 11.1  
10:00 233.6 F 8.5  
12:00 239.4 F 9.8  
14:00 247.4 F 14.0  
16:00 260.2 F 21.0  
18:00 271.5 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.4  
1:00 272.6 F 29.0  
2:00 271.2 F 28.5  
3:00 272.5 F 29.1  
4:00 271.4 F 28.4  
  
Min 271.2 F 28.3  
Max 272.8 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.3 F 0.0  
5:00 215.9 F 0.0  
10:00 227.0 F 0.0  
15:00 232.4 F 0.0  
20:00 232.0 F 0.0  
25:00 225.1 F 0.0  
30:00 222.6 F 0.0  
Drying Complete

FILLING: 2:03  
HEATING: 18:37  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:57:48  
CYCLE COMPLETE

07/24/2023 08:13 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_