
Midmark M11 - v1.0.0
Total Cycles: 4255

Name: RITTER 1
Sterilizer ID: V2178198
01/31/2024 04:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.5 F 0.0
2:00 167.4 F 3.3
4:00 185.2 F 7.5
6:00 212.7 F 12.8
8:00 225.7 F 7.3
10:00 240.5 F 12.5
12:00 236.8 F 9.5
14:00 237.3 F 8.9
16:00 248.4 F 14.3
18:00 260.4 F 21.0
20:00 271.5 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.2
1:00 272.4 F 28.7
2:00 271.3 F 28.5
3:00 272.7 F 28.9
4:00 272.9 F 29.0

Min 271.0 F 28.0
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.1 F 0.0
5:00 233.2 F 0.0
10:00 216.7 F 0.0
15:00 226.3 F 0.0
20:00 222.5 F 0.0
25:00 226.1 F 0.0
30:00 223.2 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 20:37
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 00:59:38
CYCLE COMPLETE

01/31/2024 05:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____