
Midmark M11 - v1.0.0
Total Cycles: 2772

Name: RITTER 1
Sterilizer ID: V2178198
01/12/2023 09:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.0 F 0.0
2:00 179.4 F 0.0
4:00 187.0 F 1.3
6:00 200.5 F 6.2
8:00 216.1 F 11.2
10:00 225.3 F 6.5
12:00 238.3 F 12.0
14:00 237.5 F 9.6
16:00 237.4 F 8.8
18:00 249.5 F 14.8
20:00 261.6 F 21.7
22:00 272.6 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.4 F 28.9
2:00 271.6 F 28.1
3:00 272.2 F 29.0
4:00 272.2 F 28.6

Min 271.2 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 214.7 F 0.0
10:00 228.8 F 0.0
15:00 229.4 F 0.0
20:00 226.5 F 0.0
25:00 226.3 F 0.0
30:00 227.6 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 22:26
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:25
CYCLE COMPLETE

01/12/2023 10:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____