
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002920
12:16 10/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:07
PRESSURIZING 3:11
STERILIZING 4:16
136.0 C 221kPa 4:16
136.3 C 223kPa 4:46
136.3 C 224kPa 5:16
136.1 C 219kPa 5:46
136.3 C 224kPa 6:16
136.3 C 223kPa 6:46
136.3 C 222kPa 7:16
136.3 C 224kPa 7:46
136.3 C 224kPa 7:46
VENTING 7:46
AIR DRYING 8:42
CYCLE COMPLETE 8:54

Digital Signature #
3724E8D81B65FC40
