
Midmark M11 - v1.0.0
Total Cycles: 4678

Name: RITTER 1
Sterilizer ID: V2178198
06/26/2024 10:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.5 F 0.0
2:00 160.7 F 2.6
4:00 166.6 F 5.8
6:00 190.8 F 9.6
8:00 207.0 F 10.9
10:00 220.4 F 7.2
12:00 234.7 F 11.6
14:00 231.2 F 7.8
16:00 242.8 F 13.2
18:00 241.6 F 11.5
20:00 253.8 F 17.7
22:00 264.6 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.6
1:00 272.5 F 29.3
2:00 272.3 F 29.2
3:00 272.5 F 29.7
4:00 272.2 F 29.2

Min 271.0 F 28.5
Max 273.1 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.6 F 0.0
5:00 218.3 F 0.0
10:00 226.8 F 0.0
15:00 215.4 F 0.0
20:00 224.4 F 0.0
25:00 224.9 F 0.0
30:00 227.6 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 23:58
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:02:45
CYCLE COMPLETE

06/26/2024 11:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____