

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 005200  
12:15 06/03/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:55  
PRESSURIZING 2:00  
STERILIZING 3:08  
136.4 C 220kPa 3:08  
136.3 C 223kPa 3:38  
136.4 C 224kPa 4:08  
136.4 C 224kPa 4:38  
136.4 C 225kPa 5:08  
136.2 C 223kPa 5:38  
136.2 C 224kPa 6:08  
136.4 C 224kPa 6:38  
136.4 C 224kPa 6:38  
VENTING 6:38  
AIR DRYING 7:26  
CYCLE COMPLETE 21:41

-----

Digital Signature #  
F52815378D469DE2  
-----