
Midmark M11 - v1.0.0
Total Cycles: 4161

Name: RITTER 1
Sterilizer ID: V2178198
12/27/2023 05:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.2 F 0.0
2:00 159.8 F 3.2
4:00 176.9 F 7.0
6:00 203.8 F 12.0
8:00 221.3 F 6.0
10:00 233.2 F 11.0
12:00 232.0 F 7.3
14:00 243.9 F 13.3
16:00 242.7 F 11.7
18:00 255.3 F 18.2
20:00 266.3 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.4 F 28.9
2:00 271.3 F 28.3
3:00 272.5 F 29.0
4:00 271.7 F 28.4

Min 271.3 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.9 F 0.0
5:00 235.9 F 0.0
10:00 228.9 F 0.0
15:00 226.9 F 0.0
20:00 226.6 F 0.0
25:00 228.1 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 21:36
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:00:20
CYCLE COMPLETE

12/27/2023 06:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____