
Midmark M11 - v1.0.5
Total Cycles: 2598

Name: RITTER2
Sterilizer ID: V2313228
07/04/2023 07:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 67.2 F 0.0
2:00 69.6 F 0.0
4:00 91.4 F 0.0
6:00 101.5 F 0.0
8:00 130.2 F 0.1
10:00 170.7 F 0.5
12:00 197.5 F 2.7
14:00 204.7 F 5.8
16:00 217.0 F 9.0
18:00 228.2 F 12.7
20:00 229.7 F 7.9
22:00 239.3 F 12.4
24:00 236.8 F 9.9
26:00 238.3 F 9.4
28:00 247.7 F 14.4
30:00 258.7 F 20.5
32:00 268.6 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.9
1:00 271.9 F 29.4
2:00 272.3 F 29.2
3:00 271.6 F 29.0
4:00 272.8 F 29.6

Min 271.5 F 28.7
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.0 F 0.0
5:00 192.8 F 0.0
10:00 194.9 F 0.0
15:00 197.3 F 0.0
20:00 199.2 F 0.0
25:00 201.1 F 0.0
30:00 205.5 F 0.0

Drying Complete

FILLING: 0:55
HEATING: 33:08
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:11:16
CYCLE COMPLETE

07/04/2023 08:44 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____