
Midmark M11 - v1.0.5
Total Cycles: 2427

Name: RITTER2
Sterilizer ID: V2313228
05/29/2023 11:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.2 F 0.0
2:00 199.9 F 3.9
4:00 197.5 F 9.1
6:00 218.8 F 7.6
8:00 229.8 F 10.2
10:00 233.1 F 7.9
12:00 241.0 F 11.2
14:00 246.7 F 13.9
16:00 259.2 F 20.9
18:00 270.6 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.5 F 29.6
2:00 271.7 F 29.0
3:00 271.4 F 29.2
4:00 272.5 F 29.7

Min 271.2 F 28.7
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.2 F 0.0
5:00 216.9 F 0.0
10:00 231.9 F 0.0
15:00 232.2 F 0.0
20:00 226.3 F 0.0
25:00 225.2 F 0.0
30:00 223.3 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 18:50
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:00
CYCLE COMPLETE

05/29/2023 11:59 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____