

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000642  
12:39 09/09/2021  
WRAPPED  
135 C for 10 min.  
CYCLE START 0:00  
CONDITIONING 2:08  
PRESSURIZING 4:10  
STERILIZING 5:01  
136.0 C 229kPa 5:01  
135.8 C 229kPa 5:31  
136.1 C 230kPa 6:01  
136.3 C 233kPa 6:31  
136.4 C 233kPa 7:01  
136.4 C 233kPa 7:31  
136.4 C 232kPa 8:01  
136.4 C 233kPa 8:31  
136.4 C 233kPa 9:01  
136.3 C 232kPa 9:31  
136.3 C 232kPa 10:01  
136.3 C 232kPa 10:31  
136.4 C 233kPa 11:01  
136.4 C 231kPa 11:31  
136.1 C 232kPa 12:01  
136.3 C 232kPa 12:31  
136.3 C 232kPa 13:01  
136.2 C 231kPa 13:31  
136.3 C 232kPa 14:01  
136.4 C 229kPa 14:31  
136.3 C 232kPa 15:01  
136.3 C 232kPa 15:01  
VENTING 15:01  
AIR DRYING 15:48  
CYCLE COMPLETE 30:48  
-----

Digital Signature #  
93A9E3457FB41E10  
-----