

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1194

Name: RITTER 1  
Sterilizer ID: V2178198  
09/22/2021 10:25 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.4 F 0.0  
2:00 190.8 F 2.7  
4:00 222.0 F 7.0  
6:00 222.3 F 12.3  
8:00 233.0 F 7.5  
10:00 243.7 F 12.9  
12:00 240.2 F 10.4  
14:00 236.4 F 8.7  
16:00 251.0 F 15.6  
18:00 262.6 F 22.3  
20:00 272.6 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.1  
1:00 272.7 F 28.7  
2:00 271.3 F 28.0  
3:00 272.4 F 28.9  
4:00 272.3 F 28.4  
  
Min 271.0 F 28.0  
Max 272.8 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.8 F 0.0  
5:00 233.6 F 0.0  
10:00 225.3 F 0.0  
15:00 225.5 F 0.0  
20:00 197.1 F 0.0  
Dry Cycle Aborted

FILLING: 1:51  
HEATING: 20:18  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 22:10  
TOTAL CYCLE: 00:51:40

09/22/2021 11:17 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01