
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 000866
11:32 14/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:15
STERILIZING 4:47
136.4 C 232kPa 4:47
136.4 C 233kPa 5:17
136.3 C 231kPa 5:47
136.4 C 233kPa 6:17
135.9 C 230kPa 6:47
136.4 C 231kPa 7:17
136.2 C 230kPa 7:47
136.2 C 229kPa 8:17
136.2 C 229kPa 8:17
VENTING 8:17
AIR DRYING 9:12
CYCLE COMPLETE 24:12

Digital Signature #
76E62B12CEAE6F6F
