
Midmark M11 - v1.0.5
Total Cycles: 4074

Name: RITTER2
Sterilizer ID: V2313228
11/01/2024 10:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.6 F 0.0
2:00 157.7 F 1.4
4:00 166.0 F 4.4
6:00 189.4 F 8.4
8:00 223.8 F 12.6
10:00 228.4 F 6.4
12:00 235.1 F 10.8
14:00 233.7 F 7.9
16:00 244.6 F 13.2
18:00 242.6 F 11.9
20:00 254.4 F 17.9
22:00 264.8 F 24.2
24:00 270.7 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.4
1:00 272.8 F 29.5
2:00 272.4 F 29.9
3:00 271.5 F 29.5
4:00 272.7 F 29.6

Min 271.5 F 29.2
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 169.0 F 0.0
5:00 247.4 F 0.0
10:00 191.3 F 0.0
15:00 189.0 F 0.0
20:00 185.4 F 0.0
25:00 182.3 F 0.0
30:00 180.0 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 24:46
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:03:57
CYCLE COMPLETE

11/01/2024 11:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____