
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001609
10:43 20/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:13
STERILIZING 4:23
136.4 C 225kPa 4:23
136.4 C 225kPa 4:53
136.4 C 226kPa 5:23
136.4 C 225kPa 5:53
136.4 C 226kPa 6:23
136.3 C 225kPa 6:53
136.4 C 225kPa 7:23
136.4 C 224kPa 7:53
136.4 C 224kPa 7:53
VENTING 7:53
AIR DRYING 8:47
CYCLE COMPLETE 9:11

Digital Signature #
DB3575D34537A517
