

-----  
Midmark M11 - v1.0.0  
Total Cycles: 971

Name: RITTER 1  
Sterilizer ID: V2178198  
07/06/2021 03:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 199.3 F 0.0  
2:00 193.9 F 3.9  
4:00 209.1 F 8.9  
6:00 229.8 F 9.0  
8:00 231.7 F 9.9  
10:00 236.1 F 8.5  
12:00 245.9 F 13.1  
14:00 244.6 F 12.4  
16:00 257.3 F 19.3  
18:00 268.7 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.4 F 28.9  
2:00 272.8 F 29.3  
3:00 271.4 F 28.7  
4:00 272.8 F 29.1  
  
Min 271.4 F 28.6  
Max 273.1 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.2 F 0.0  
5:00 233.5 F 0.0  
10:00 230.5 F 0.0  
15:00 232.5 F 0.0  
20:00 226.5 F 0.0  
25:00 225.3 F 0.0  
30:00 181.3 F 0.0  
Drying Complete

FILLING: 2:23  
HEATING: 19:50  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:59:27  
CYCLE COMPLETE

07/06/2021 04:00 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02