
Midmark M11 - v1.0.5
Total Cycles: 2851

Name: RITTER2
Sterilizer ID: V2313228
09/08/2023 10:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.2 F 0.0
2:00 183.6 F 3.0
4:00 191.4 F 7.7
6:00 211.1 F 12.8
8:00 222.4 F 7.9
10:00 236.1 F 13.1
12:00 237.5 F 10.8
14:00 236.6 F 9.5
16:00 250.2 F 16.1
18:00 261.5 F 22.5
20:00 271.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.7 F 29.8
2:00 271.3 F 29.0
3:00 272.3 F 30.0
4:00 272.2 F 29.5

Min 271.1 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.2 F 0.0
5:00 214.3 F 0.0
10:00 232.9 F 0.0
15:00 231.4 F 0.0
20:00 224.3 F 0.0
25:00 223.8 F 0.0
30:00 226.8 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 20:36
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 00:59:56
CYCLE COMPLETE

09/08/2023 11:20 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____