
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002242
13:29 11/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:10
PRESSURIZING 2:16
STERILIZING 3:21
136.5 C 224kPa 3:21
136.2 C 224kPa 3:51
136.3 C 225kPa 4:21
136.4 C 222kPa 4:51
136.2 C 222kPa 5:21
136.4 C 226kPa 5:51
136.4 C 225kPa 6:21
136.4 C 223kPa 6:51
136.4 C 223kPa 6:51
VENTING 6:52
AIR DRYING 7:43
CYCLE COMPLETE 8:15

Digital Signature #
64B64E7D2990F184
