

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2096

Name: RITTER 1  
Sterilizer ID: V2178198  
07/07/2022 07:15 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.5 F 0.0  
2:00 74.8 F 0.0  
4:00 85.2 F 0.0  
6:00 120.8 F 1.0  
8:00 151.0 F 2.7  
10:00 175.9 F 5.2  
12:00 194.3 F 8.5  
14:00 213.1 F 12.1  
16:00 219.7 F 4.5  
18:00 229.5 F 8.5  
20:00 238.1 F 12.5  
22:00 235.6 F 8.9  
24:00 241.8 F 10.6  
26:00 244.9 F 12.6  
28:00 256.5 F 18.8  
30:00 267.1 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.7 F 29.1  
2:00 271.7 F 28.2  
3:00 271.9 F 28.9  
4:00 272.4 F 28.7  
  
Min 271.1 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.2 F 0.0  
5:00 221.6 F 0.0  
10:00 199.8 F 0.0  
15:00 191.1 F 0.0  
20:00 190.4 F 0.0  
25:00 192.6 F 0.0  
30:00 194.5 F 0.0  
Drying Complete

FILLING: 0:36  
HEATING: 31:35  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:09:17  
CYCLE COMPLETE

07/07/2022 08:24 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_