
Midmark M11 - v1.0.0
Total Cycles: 1338

Name: RITTER 1
Sterilizer ID: V2178198
11/08/2021 09:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 106.2 F 0.0
2:00 107.0 F 0.0
4:00 117.2 F 0.1
6:00 153.4 F 0.2
8:00 195.5 F 1.8
10:00 213.2 F 5.2
12:00 224.5 F 8.5
14:00 234.3 F 12.4
16:00 231.1 F 7.6
18:00 241.5 F 12.5
20:00 238.1 F 9.8
22:00 239.1 F 9.1
24:00 248.6 F 14.3
26:00 260.0 F 20.6
28:00 270.4 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.6 F 28.8
2:00 271.5 F 27.9
3:00 272.5 F 28.9
4:00 272.1 F 28.3

Min 271.4 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 227.3 F 0.0
10:00 225.6 F 0.0
15:00 211.9 F 0.0
20:00 195.2 F 0.0
25:00 195.8 F 0.0
30:00 176.4 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 28:52
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:07:03
CYCLE COMPLETE

11/08/2021 10:57 AM
ACCEPTED: X
REJECTED:
Operator: U01