
Midmark M11 - v1.0.0
Total Cycles: 2513

Name: RITTER 1
Sterilizer ID: V2178198
11/03/2022 03:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.9 F 0.0
2:00 199.6 F 4.2
4:00 207.1 F 9.2
6:00 225.7 F 6.6
8:00 232.3 F 10.3
10:00 234.6 F 7.5
12:00 244.0 F 12.4
14:00 244.9 F 12.6
16:00 257.7 F 19.5
18:00 269.1 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 272.2 F 28.9
2:00 271.3 F 28.2
3:00 272.1 F 29.0
4:00 271.9 F 28.6

Min 271.1 F 28.2
Max 272.3 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.9 F 0.0
5:00 242.2 F 0.0
10:00 231.2 F 0.0
15:00 229.3 F 0.0
20:00 229.8 F 0.0
25:00 228.6 F 0.0
30:00 229.2 F 0.0

Drying Complete

FILLING: 1:57
HEATING: 19:05
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:13
CYCLE COMPLETE

11/03/2022 04:10 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____