

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 005917  
9:46 06/07/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:34  
PRESSURIZING 2:38  
STERILIZING 3:50  
136.5 C 222kPa 3:50  
136.4 C 225kPa 4:20  
136.2 C 223kPa 4:50  
136.4 C 223kPa 5:20  
136.4 C 225kPa 5:50  
136.2 C 224kPa 6:20  
136.3 C 219kPa 6:50  
136.3 C 223kPa 7:20  
136.3 C 223kPa 7:20  
VENTING 7:20  
AIR DRYING 8:08  
CYCLE COMPLETE 8:45

-----  
Digital Signature #  
91737193B64F6664  
-----