

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4806

Name: RITTER 1  
Sterilizer ID: V2178198  
08/12/2024 04:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 175.5 F 0.0  
2:00 175.3 F 2.8  
4:00 197.7 F 7.5  
6:00 211.8 F 12.5  
8:00 223.6 F 7.3  
10:00 237.7 F 12.8  
12:00 237.3 F 10.3  
14:00 235.4 F 8.2  
16:00 249.7 F 15.3  
18:00 261.8 F 22.0  
20:00 272.3 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.5  
1:00 272.4 F 29.1  
2:00 271.2 F 28.4  
3:00 272.5 F 29.3  
4:00 271.8 F 28.7  
  
Min 271.2 F 28.4  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.3 F 0.0  
5:00 216.5 F 0.0  
10:00 223.7 F 0.0  
15:00 221.9 F 0.0  
20:00 223.0 F 0.0  
25:00 229.5 F 0.0  
30:00 228.4 F 0.0  
Drying Complete

FILLING: 1:43  
HEATING: 20:29  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 00:59:44  
CYCLE COMPLETE

08/12/2024 05:00 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_