

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3298

Name: RITTER 1  
Sterilizer ID: V2178198  
05/29/2023 01:08 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 211.5 F 0.0  
2:00 202.7 F 2.5  
4:00 225.3 F 8.3  
6:00 233.8 F 10.6  
8:00 237.6 F 10.6  
10:00 236.5 F 8.3  
12:00 240.5 F 10.2  
14:00 248.9 F 14.4  
16:00 261.2 F 21.3  
18:00 272.4 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.0  
1:00 272.0 F 28.7  
2:00 272.1 F 28.4  
3:00 271.3 F 28.0  
4:00 272.4 F 28.9  
  
Min 271.3 F 27.9  
Max 272.5 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.2 F 0.0  
5:00 230.6 F 0.0  
10:00 228.1 F 0.0  
15:00 229.5 F 0.0  
20:00 226.4 F 0.0  
25:00 231.0 F 0.0  
30:00 225.5 F 0.0  
Drying Complete

FILLING: 2:33  
HEATING: 18:21  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:58:12  
CYCLE COMPLETE

05/29/2023 02:06 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_