

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4904

Name: RITTER 1  
Sterilizer ID: V2178198  
09/12/2024 06:06 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 147.9 F 0.0  
2:00 159.8 F 0.3  
4:00 190.8 F 3.2  
6:00 213.0 F 8.7  
8:00 229.1 F 8.5  
10:00 238.1 F 10.5  
12:00 233.5 F 7.8  
14:00 240.5 F 10.2  
16:00 247.9 F 13.9  
18:00 261.1 F 21.2  
20:00 272.7 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 27.9  
1:00 272.6 F 28.7  
2:00 271.5 F 27.8  
3:00 272.5 F 28.6  
4:00 271.5 F 27.8  
  
Min 271.2 F 27.8  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.0 F 0.0  
5:00 240.1 F 0.0  
10:00 235.7 F 0.0  
15:00 229.0 F 0.0  
20:00 216.3 F 0.0  
25:00 193.8 F 0.0  
30:00 182.7 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 20:13  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:36  
CYCLE COMPLETE

09/12/2024 07:04 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_