

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2287

Name: RITTER2  
Sterilizer ID: V2313228  
04/21/2023 05:18 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 183.6 F 0.0  
2:00 183.2 F 4.1  
4:00 200.9 F 9.7  
6:00 218.1 F 5.6  
8:00 234.4 F 11.1  
10:00 236.0 F 9.4  
12:00 237.4 F 9.0  
14:00 252.3 F 16.7  
16:00 265.0 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.7  
1:00 272.4 F 28.9  
2:00 272.9 F 29.5  
3:00 271.7 F 28.6  
4:00 272.9 F 29.3  
  
Min 271.7 F 28.5  
Max 273.0 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.5 F 0.0  
5:00 208.1 F 0.0  
10:00 222.0 F 0.0  
15:00 217.0 F 0.0  
20:00 211.3 F 0.0  
25:00 208.7 F 0.0  
30:00 205.8 F 0.0  
  
Drying Complete

FILLING: 2:37  
HEATING: 17:37  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 30:00  
TOTAL CYCLE: 00:57:17  
CYCLE COMPLETE

04/21/2023 06:15 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_