
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000053
9:38 28/09/2020
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 6:25
PRESSURIZING 8:47
STERILIZING 10:57
136.0 C 222kPa 10:57
136.1 C 223kPa 11:27
136.5 C 227kPa 11:57
136.4 C 223kPa 12:27
136.2 C 225kPa 12:57
136.3 C 225kPa 13:27
136.3 C 225kPa 13:57
136.4 C 226kPa 14:27
VENTING 14:27
AIR DRYING 15:44
CYCLE COMPLETE 45:44

Digital Signature #
CABFBFFF0F5740A6
