
Midmark M11 - v1.0.0
Total Cycles: 3474

Name: RITTER 1
Sterilizer ID: V2178198
06/30/2023 01:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.3 F 0.0
2:00 186.4 F 4.0
4:00 213.1 F 9.2
6:00 225.2 F 7.2
8:00 230.8 F 10.5
10:00 234.4 F 7.5
12:00 242.9 F 12.0
14:00 245.3 F 12.9
16:00 257.8 F 19.5
18:00 268.8 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.1 F 28.6
2:00 272.1 F 28.9
3:00 271.7 F 28.3
4:00 272.5 F 29.0

Min 271.2 F 28.2
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 213.1 F 0.0
10:00 217.6 F 0.0
15:00 218.3 F 0.0
20:00 218.6 F 0.0
25:00 214.8 F 0.0
30:00 212.7 F 0.0
Drying Complete

FILLING: 2:02
HEATING: 19:04
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:58:27
CYCLE COMPLETE

06/30/2023 02:21 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____