

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3799

Name: RITTER 1  
Sterilizer ID: V2178198  
09/20/2023 10:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.0 F 0.0  
2:00 181.4 F 0.1  
4:00 186.8 F 2.9  
6:00 207.3 F 7.6  
8:00 230.2 F 12.8  
10:00 233.9 F 8.4  
12:00 242.7 F 12.2  
14:00 242.6 F 11.5  
16:00 239.8 F 10.0  
18:00 252.9 F 16.5  
20:00 264.1 F 23.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 27.9  
1:00 272.1 F 28.6  
2:00 271.8 F 28.2  
3:00 272.4 F 28.6  
4:00 272.5 F 28.9

Min 271.0 F 27.8  
Max 272.8 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.4 F 0.0  
5:00 216.8 F 0.0  
10:00 200.9 F 0.0  
15:00 198.7 F 0.0  
20:00 193.4 F 0.0  
25:00 189.0 F 0.0  
30:00 185.0 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 21:56  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:00:26  
CYCLE COMPLETE

09/20/2023 11:47 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_