
Midmark M11 - v1.0.5
Total Cycles: 2938

Name: RITTER2
Sterilizer ID: V2313228
09/27/2023 12:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.3 F 0.0
2:00 154.1 F 3.2
4:00 169.8 F 6.9
6:00 194.1 F 12.0
8:00 214.9 F 6.1
10:00 230.5 F 11.4
12:00 232.1 F 8.4
14:00 241.2 F 11.1
16:00 245.4 F 13.3
18:00 257.8 F 20.0
20:00 268.7 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.6 F 29.5
2:00 271.4 F 28.7
3:00 272.4 F 29.7
4:00 272.3 F 29.2

Min 271.2 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.2 F 0.0
5:00 207.5 F 0.0
10:00 210.4 F 0.0
15:00 200.0 F 0.0
20:00 195.4 F 0.0
25:00 196.6 F 0.0
30:00 207.0 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 21:10
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:34
CYCLE COMPLETE

09/27/2023 01:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____