
Midmark M11 - v1.0.5
Total Cycles: 639

Name: RITTER2
Sterilizer ID: V2313228
12/15/2021 03:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.2 F 0.0
2:00 184.4 F 3.4
4:00 196.6 F 8.4
6:00 214.9 F 12.2
8:00 225.7 F 9.1
10:00 234.2 F 10.6
12:00 240.7 F 12.3
14:00 241.3 F 11.6
16:00 254.3 F 18.4
18:00 265.8 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.5 F 30.0
2:00 271.8 F 29.4
3:00 271.5 F 29.6
4:00 272.4 F 30.1

Min 271.3 F 29.1
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.1 F 0.0
5:00 217.1 F 0.0
10:00 221.7 F 0.0
15:00 220.0 F 0.0
20:00 215.8 F 0.0
25:00 214.8 F 0.0
30:00 213.0 F 0.0
Drying Complete

FILLING: 2:39
HEATING: 19:43
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:39
CYCLE COMPLETE

12/16/2021 07:44 AM
ACCEPTED: X
REJECTED:
Operator: U02