

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2903

Name: RITTER 1  
Sterilizer ID: V2178198  
02/14/2023 06:21 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.5 F 0.0  
2:00 182.8 F 4.1  
4:00 200.3 F 8.7  
6:00 219.8 F 9.5  
8:00 226.8 F 8.7  
10:00 235.5 F 10.4  
12:00 239.4 F 11.1  
14:00 237.3 F 9.2  
16:00 250.4 F 15.7  
18:00 261.8 F 22.2  
20:00 272.5 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.7  
1:00 272.2 F 29.2  
2:00 271.3 F 28.7  
3:00 272.3 F 29.3  
4:00 271.5 F 28.7  
  
Min 271.3 F 28.6  
Max 272.5 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.4 F 0.0  
5:00 225.4 F 0.0  
10:00 228.3 F 0.0  
15:00 237.9 F 0.0  
20:00 229.8 F 0.0  
25:00 229.2 F 0.0  
30:00 227.3 F 0.0  
Drying Complete

FILLING: 1:38  
HEATING: 20:27  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:59:21  
CYCLE COMPLETE

02/14/2023 07:20 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_