

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2976

Name: RITTER2  
Sterilizer ID: V2313228  
10/05/2023 09:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 152.2 F 0.0  
2:00 147.3 F 1.3  
4:00 152.6 F 4.2  
6:00 177.2 F 8.2  
8:00 199.9 F 13.0  
10:00 219.5 F 6.7  
12:00 232.4 F 11.7  
14:00 233.2 F 8.9  
16:00 239.8 F 10.5  
18:00 245.8 F 13.6  
20:00 258.2 F 20.3  
22:00 269.1 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 272.2 F 29.1  
1:00 272.4 F 29.3  
2:00 272.6 F 29.6  
3:00 272.1 F 29.7  
4:00 271.5 F 29.0

Min 271.5 F 28.9  
Max 272.6 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.5 F 0.0  
5:00 232.0 F 0.0  
10:00 216.1 F 0.0  
15:00 215.0 F 0.0  
20:00 210.3 F 0.0  
25:00 207.3 F 0.0  
30:00 203.3 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 22:51  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:01:13  
CYCLE COMPLETE

10/05/2023 10:08 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_