
Midmark M11 - v1.0.5
Total Cycles: 501

Name: RITTER2
Sterilizer ID: V2313228
10/29/2021 06:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.4 F 0.0
2:00 76.8 F 0.0
4:00 79.1 F 0.0
6:00 108.1 F 0.2
8:00 158.5 F 0.3
10:00 181.0 F 2.4
12:00 212.2 F 6.1
14:00 222.4 F 9.9
16:00 225.0 F 9.9
18:00 232.1 F 8.8
20:00 242.9 F 12.9
22:00 241.2 F 11.5
24:00 239.2 F 10.3
26:00 252.4 F 16.9
28:00 263.8 F 23.6
30:00 271.6 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.6 F 29.0
2:00 272.3 F 29.5
3:00 271.8 F 28.6
4:00 272.4 F 29.3

Min 271.5 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.4 F 0.0
5:00 237.6 F 0.0
10:00 234.3 F 0.0
15:00 222.0 F 0.0
20:00 211.6 F 0.0
25:00 217.1 F 0.0
30:00 213.7 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 30:03
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 01:08:02
CYCLE COMPLETE

10/29/2021 08:49 AM
ACCEPTED: X
REJECTED:
Operator: U02