
Midmark M11 - v1.0.0
Total Cycles: 4662

Name: RITTER 1
Sterilizer ID: V2178198
06/19/2024 04:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.7 F 0.0
2:00 185.5 F 3.6
4:00 205.2 F 7.8
6:00 213.5 F 12.5
8:00 227.1 F 6.6
10:00 234.1 F 11.3
12:00 234.1 F 8.1
14:00 244.0 F 12.1
16:00 245.1 F 12.6
18:00 256.8 F 18.9
20:00 267.7 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 272.5 F 28.7
2:00 271.1 F 28.0
3:00 272.6 F 28.9
4:00 272.1 F 28.9

Min 271.0 F 27.9
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.3 F 0.0
5:00 231.5 F 0.0
10:00 223.0 F 0.0
15:00 200.5 F 0.0
20:00 185.2 F 0.0
25:00 192.9 F 0.0
30:00 201.5 F 0.0
Drying Complete

FILLING: 1:53
HEATING: 21:24
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:01:06
CYCLE COMPLETE

06/19/2024 05:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____