

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 001548  
10:22 30/10/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:41  
PRESSURIZING 2:48  
STERILIZING 3:49  
136.4 C 221kPa 3:49  
136.0 C 221kPa 4:19  
136.3 C 221kPa 4:49  
136.1 C 224kPa 5:19  
136.3 C 224kPa 5:49  
136.4 C 224kPa 6:19  
136.0 C 221kPa 6:49  
136.2 C 222kPa 7:19  
136.2 C 222kPa 7:19  
VENTING 7:19  
AIR DRYING 8:09  
CYCLE COMPLETE 8:19  
-----

Digital Signature #  
9004D42D7AAAA14E  
-----