

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001396  
9:33 25/03/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:34  
PRESSURIZING 2:38  
STERILIZING 3:38  
136.0 C 220kPa 3:38  
136.1 C 221kPa 4:08  
136.4 C 224kPa 4:38  
136.1 C 219kPa 5:08  
136.4 C 223kPa 5:38  
136.3 C 223kPa 6:08  
136.2 C 223kPa 6:38  
136.3 C 220kPa 7:08  
136.3 C 220kPa 7:08  
VENTING 7:08  
AIR DRYING 8:01  
CYCLE COMPLETE 8:05

-----  
Digital Signature #  
DDB8E8CA2E0B72DD  
-----