
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002037
11:52 13/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:55
PRESSURIZING 2:00
STERILIZING 3:02
136.0 C 222kPa 3:02
136.2 C 224kPa 3:32
136.4 C 225kPa 4:02
136.3 C 225kPa 4:32
136.3 C 223kPa 5:02
136.2 C 224kPa 5:32
136.3 C 225kPa 6:02
136.4 C 223kPa 6:32
136.4 C 223kPa 6:32
VENTING 6:32
AIR DRYING 7:24
CYCLE COMPLETE 37:24

Digital Signature #
C32F01CA5436562E
