
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004057
14:28 19/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:37
PRESSURIZING 3:38
STERILIZING 4:42
136.4 C 225kPa 4:42
136.4 C 225kPa 5:12
136.4 C 223kPa 5:42
136.4 C 224kPa 6:12
136.4 C 224kPa 6:42
136.1 C 223kPa 7:12
136.4 C 222kPa 7:42
136.4 C 225kPa 8:12
136.4 C 225kPa 8:12
VENTING 8:12
AIR DRYING 9:03
CYCLE COMPLETE 9:43

Digital Signature #
3FAFBD5D59C63AB2
