
Midmark M11 - v1.0.0
Total Cycles: 3922

Name: RITTER 1
Sterilizer ID: V2178198
10/19/2023 09:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.7 F 0.0
2:00 165.6 F 1.9
4:00 166.0 F 5.4
6:00 188.0 F 9.7
8:00 219.8 F 6.0
10:00 222.6 F 8.2
12:00 232.8 F 13.4
14:00 236.5 F 10.1
16:00 236.5 F 8.5
18:00 247.9 F 14.3
20:00 259.4 F 20.6
22:00 269.9 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.2 F 29.1
2:00 272.9 F 29.1
3:00 272.0 F 28.7
4:00 271.4 F 28.4

Min 271.1 F 28.3
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 225.6 F 0.0
10:00 228.5 F 0.0
15:00 230.1 F 0.0
20:00 231.4 F 0.0
25:00 229.8 F 0.0
30:00 227.0 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 23:05
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:01:38
CYCLE COMPLETE

10/19/2023 10:27 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____