
Midmark M11 - v1.0.0
Total Cycles: 2938

Name: RITTER 1
Sterilizer ID: V2178198
02/23/2023 12:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.5 F 0.0
2:00 182.2 F 4.6
4:00 192.6 F 9.1
6:00 210.8 F 7.6
8:00 223.5 F 8.6
10:00 236.0 F 11.0
12:00 239.2 F 11.1
14:00 237.3 F 9.4
16:00 251.8 F 16.3
18:00 263.8 F 23.4
20:00 271.4 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.6 F 29.1
2:00 271.5 F 28.4
3:00 272.7 F 29.3
4:00 272.0 F 28.7

Min 271.4 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.0 F 0.0
5:00 227.4 F 0.0
10:00 240.3 F 0.0
15:00 229.9 F 0.0
20:00 229.4 F 0.0
25:00 230.9 F 0.0
30:00 228.0 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 20:00
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:31
CYCLE COMPLETE

02/23/2023 01:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____