
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000906
12:29 04/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:58
PRESSURIZING 2:05
STERILIZING 3:14
135.9 C 220kPa 3:14
136.4 C 233kPa 3:44
136.2 C 230kPa 4:14
136.3 C 233kPa 4:44
136.1 C 230kPa 5:14
136.3 C 233kPa 5:44
136.1 C 230kPa 6:14
136.4 C 233kPa 6:44
136.4 C 233kPa 6:44
VENTING 6:44
AIR DRYING 7:37
CYCLE COMPLETE 22:37

Digital Signature #
42A68290A1066002
