
Midmark M11 - v1.0.0
Total Cycles: 1098

Name: RITTER 1
Sterilizer ID: V2178198
08/19/2021 09:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

C232: FILL MODE
WATER LOW
CHAMBER TEMP 200F
CHAMBER PRESS 0.0PSI
** ITEMS NOT STERILE **

08/19/2021 09:52 AM