

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3242

Name: RITTER 1  
Sterilizer ID: V2178198  
05/16/2023 12:42 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 175.7 F 0.0  
2:00 174.5 F 0.3  
4:00 208.3 F 5.0  
6:00 222.5 F 10.1  
8:00 226.9 F 6.7  
10:00 237.9 F 11.0  
12:00 234.3 F 8.4  
14:00 240.6 F 10.5  
16:00 246.9 F 13.8  
18:00 259.3 F 20.5  
20:00 270.4 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.3  
1:00 272.0 F 28.9  
2:00 272.2 F 28.7  
3:00 271.2 F 28.3  
4:00 272.6 F 29.1

Min 271.2 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.7 F 0.0  
5:00 243.2 F 0.0  
10:00 244.4 F 0.0  
15:00 231.7 F 0.0  
20:00 225.6 F 0.0  
25:00 220.8 F 0.0  
30:00 218.8 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 20:43  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 00:58:59  
CYCLE COMPLETE

05/16/2023 01:40 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_