
Midmark M11 - v1.0.0
Total Cycles: 5124

Name: RITTER 1
Sterilizer ID: V2178198
11/21/2024 06:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.8 F 0.0
2:00 177.7 F 0.4
4:00 205.1 F 3.2
6:00 220.3 F 8.2
8:00 236.8 F 13.2
10:00 235.3 F 9.6
12:00 236.4 F 8.8
14:00 244.3 F 12.0
16:00 246.6 F 13.4
18:00 260.0 F 20.8
20:00 272.0 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.4 F 28.5
2:00 272.3 F 28.8
3:00 271.9 F 28.2
4:00 272.7 F 29.0

Min 271.5 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.4 F 0.0
5:00 197.7 F 0.0
10:00 205.0 F 0.0
15:00 205.4 F 0.0
20:00 202.7 F 0.0
25:00 214.1 F 0.0
30:00 246.0 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 20:31
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 00:59:10
CYCLE COMPLETE

11/21/2024 07:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____