
Midmark M11 - v1.0.0
Total Cycles: 1832

Name: RITTER 1
Sterilizer ID: V2178198
04/18/2022 09:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.3 F 0.0
2:00 191.1 F 3.3
4:00 189.5 F 8.3
6:00 206.9 F 13.5
8:00 227.5 F 7.5
10:00 240.4 F 12.5
12:00 236.7 F 9.4
14:00 239.7 F 9.3
16:00 248.0 F 13.8
18:00 259.9 F 20.4
20:00 270.5 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 271.9 F 28.3
2:00 272.4 F 28.5
3:00 272.5 F 28.7
4:00 272.1 F 28.9

Min 271.5 F 28.3
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.3 F 0.0
5:00 234.1 F 0.0
10:00 230.6 F 0.0
15:00 229.3 F 0.0
20:00 221.5 F 0.0
25:00 228.5 F 0.0
30:00 234.1 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 21:29
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:43
CYCLE COMPLETE

04/18/2022 10:39 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____