
Midmark M11 - v1.0.0
Total Cycles: 3296

Name: RITTER 1
Sterilizer ID: V2178198
05/29/2023 11:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.7 F 0.0
2:00 192.4 F 1.1
4:00 208.0 F 6.6
6:00 223.8 F 12.1
8:00 230.7 F 8.3
10:00 240.3 F 11.1
12:00 243.8 F 12.4
14:00 242.4 F 11.3
16:00 255.9 F 18.3
18:00 267.6 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.5 F 28.8
2:00 271.5 F 28.1
3:00 272.2 F 28.9
4:00 272.2 F 28.6

Min 271.3 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.5 F 0.0
5:00 233.6 F 0.0
10:00 241.7 F 0.0
15:00 240.0 F 0.0
20:00 227.3 F 0.0
25:00 222.9 F 0.0
30:00 221.5 F 0.0
Drying Complete

FILLING: 2:02
HEATING: 19:21
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:40
CYCLE COMPLETE

05/29/2023 12:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____