
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000982
11:13 10/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:09
PRESSURIZING 2:15
STERILIZING 3:17
136.4 C 233kPa 3:17
136.4 C 233kPa 3:47
136.4 C 232kPa 4:17
136.3 C 233kPa 4:47
136.3 C 233kPa 5:17
136.2 C 232kPa 5:47
136.4 C 233kPa 6:17
136.4 C 233kPa 6:47
136.4 C 233kPa 6:47
VENTING 6:47
AIR DRYING 7:37
CYCLE COMPLETE 8:06

Digital Signature #
EBD6420A344D2A0C
