
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001615
8:47 11/25/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:04
PRESSURIZING 6:57
STERILIZING 8:59
136.0 C 220kPa 8:59
136.3 C 220kPa 9:29
136.2 C 219kPa 9:59
136.4 C 220kPa 10:30
136.5 C 222kPa 11:00
136.2 C 220kPa 11:30
136.3 C 221kPa 12:00
136.0 C 218kPa 12:29
VENTING 12:29
AIR DRYING 13:31
CYCLE COMPLETE 42:05

Digital Signature #
94FB66FB7035A8DD
